

ANNEXURE

List of medical equipment not provided/not maintained/not operated

Sl. No	Name of equipment/ Machine	Status
North Eastern Railway		
1.	TMT Machine (cost Rs.0.08 crore)	Commissioned in 1992 but used only for one day in 2003-04.
2.	Boyle's Apparatus used with Ventilator	The Apparatus has only one sensor against the requirement of two sensors. Procurement proposal for the second censor was mooted only in 2004-05.
3.	Multi Parameter Monitor with companometer	Equipment is used to monitor pulse rate, heart, blood pressure and electrocardiogram during surgery. The equipment had not been procured despite instructions (April 2000) from Director General, Health Services.
4.	Boyles Apparatus Anaesthesia Machine with multi monitors (BAAM)	Old BAAM, which was purchased 20 years ago, had outlived its life and broke down frequently. Sanction for its replacement had not been granted causing risk to life of those patients undergoing surgeries.
5.	Bird Mark Ventilators (2)	Equipment essential for supply of oxygen to patients in ICUs are out of order since 2003. Replacement had not been sanctioned so far.
6.	Computerised X-Ray Machine(CXR)	Proposal for procurement of this machine capable of taking clear images for better diagnosis, reducing the radiation, quicker and cost effective was sent in February 2003 but had not been sanctioned (September 2005).
7.	Blood Gas Analyser	Equipment installed in 1993 went out of order within two years due leakage in gas cylinder. Despite approval for replacement in November 2001, the new equipment had not been installed so far (September 2005).
East Coast Railway		
1.	Semi-auto Analyser	This equipment, purchased in 1995, is lying in break down condition since 2003.
2.	Flame Photometer	This equipment, purchased in 1999, is lying in break down condition since 2003.
3.	Portable 60 MA X-ray Machine	This machine, procured in 1987, has been out of order for the last 10 years. The equipment is under condemnation.
South East Central Railway		
1.	Portable 60 MA X-Ray Machine	The machine, installed in July 1980, had gone out of order. Non-replacement of the machine till December 2004 resulted in the work (12245 X-rays) being got done at higher cost from outside.
Northeast Frontier Railway		
1.	Hemeodialysis Machine	The machine is lying idle since its installation in July 2001 as no operator has been provided to work on this machine (cost Rs.0.08 crore).
2.	Reserve Osmosis water treatment plant used for Hemeodialysis	The plant has not been used since its installation in December 2001(cost Rs.0.05 crore)
3.	Vinitic spectra photometer -Auto analyser Elisa Processor	The equipment has not been used since installation in May 2001 (cost Rs.0.02 crore).
4.	1 st Track Ultrasound scanner	These machines were idling (since January 2004) for want of posting of a radiologist (cost Rs.0.17 crore)
5.	300 MAX-ray Machine	